

EBV - G (VCA) ANTIBODY TEST SYSTEM



Bion EBV-G KITS AND REAGENTS

KITS and KIT COMPONENTS

CODE NO.

EBV-G (EBV-VCA IgG Antibody) 120 Test Kit	EBG-120
EBV-VCA Substrate Slide, twelve wells	EB-5012
EBV-VCA IgG Positive Control Serum, 0.5 ml	EBG-5020
EBV-VCA Negative Control Serum, 0.5 ml	EBN-5010
Conjugate, IgG with Counterstain, 3.5 ml	CXCG-9972
Mounting Medium, 3.5 ml	MM-9985
PBS Packet, One Liter	PBS-9990

NOTE: Changes highlighted

SYMBOL DEFINITIONS

= Consult Directions for Use	= In Vitro Diagnostic Reagent
= Store Away From Direct Light	= Positive Control
= Storage Temperature	= Negative Control
= Expiration Date	= Endpoint Titer
= Number of Tests	= Code Number
= Amount	= Lot Number
= Single Use Only	= Contains biological material of animal origin
= Contains human blood or plasma derivatives	

INTENDED USE

The Bion EBV-G (VCA) ANTIBODY TEST SYSTEM (Epstein-Barr Virus Viral Capsid Antigen) is an indirect fluorescent antibody assay for the qualitative and/or semi-quantitative determination of EBV-VCA IgG antibodies in human serum. The Bion EBV-G (VCA) ANTIBODY TEST SYSTEM is intended for use as an aid in the diagnosis of primary infection (Infectious Mononucleosis), reinfection or reactivation of the latent virus and as a determination of immunological experience with EBV.

SUMMARY AND EXPLANATION

The Epstein-Barr Virus (EBV) was first detected and described by Epstein, Barr and Achong in electron-microscopic studies of lymphoblastoid cultures obtained from patients with Burkitt's Lymphoma.^{1,2} The etiologic role of EBV in Infectious Mononucleosis (IM) was first reported in 1968 and has since been well documented to be the causative agent.^{3,4,5} Infection with EBV results in the expression of several viral proteins to which the host responds with appropriate antibodies. One of these viral proteins is the Viral Capsid Antigen (VCA).

Although the heterophile antibody response is relatively specific for IM, it has been observed that most children and up to 10% of adult patients fail to develop heterophile antibodies.⁶ Cytomegalovirus and Toxoplasma gondii also produce infectious mononucleosis-like disease which is often clinically indistinguishable from EBV Infectious Mononucleosis. Testing for EBV specific antibodies should be performed in all cases of infectious mononucleosis-like disease and especially those that are heterophile negative.

Elevated antibody titers to EBV-VCA develop in patients with Burkitt's Lymphoma, Nasopharyngeal Carcinoma, and EBV Infectious Mononucleosis. A high incidence of elevated titers is also found in Hodgkin's Disease, Lymphocytic Leukemia, Systemic Lupus Erythematosus, Sarcoidosis, and Izumi Fever.^{7,8,9}

Many symptomatic patients have already reached peak titers of IgG to EBV-VCA by the time they consult their physicians, and further increases cannot be demonstrated.¹⁰ Testing for IgM specific EBV-VCA antibodies should be done in these cases. Methods for EBV-VCA antibody detection include Enzyme-Linked Immunosorbent Assay (ELISA) and Immunofluorescence Assay (IFA).¹¹ The ELISA method is not able to dependably differentiate between specific and nonspecific reactions. However, the localization of the antigen-antibody reaction can be visualized with the IFA test, thus allowing the specific antigen-antibody reaction to be recognized by the morphology of the fluorescence staining.¹¹

PRINCIPLE OF THE IFA PROCEDURE

The Bion EBV-G VCA ANTIBODY TEST SYSTEM utilizes the indirect fluorescent antibody assay method first described by Weller and Coons¹² and further developed by Riggs, et al.¹³ The procedure is carried out in two basic reaction steps:

Step 1 - Human serum is reacted with the antigen substrate. Antibodies, if present, will bind to the antigen forming stable antigen-antibody complexes. If no antibodies are present, the complexes will not be formed and serum components will be washed away.

Step 2 - Fluorescein labeled antihuman IgG antibody is added to the reaction site which binds with the complexes formed in step one. This results in a positive reaction of bright apple-green fluorescence when viewed with a properly equipped fluorescence microscope. If no complexes are formed in step one, the fluorescein labeled antibody will be washed away, exhibiting a negative result.

REAGENTS

EBV-VCA ANTIGEN SUBSTRATE SLIDES

Ten individually foil-wrapped twelve well slides with a mixture of Epstein-Barr Virus (P3HR1 strain) infected and uninfected lymphocytic cells fixed onto each well. Each well contains an average of 10% infected cells per 200X field. Stable in sealed foil pouch at 8°C, or lower, until labeled expiration date.

POSITIVE CONTROL SERUM

One vial containing 0.5 ml EBV-VCA positive IgG human control serum with protein stabilizer and 0.005% thimerosal. Stable at 2-8°C until labeled expiration date.

When used undiluted as provided, specific fluorescent intensity of 3+ or greater should be seen. Optionally, the positive control can be titrated to endpoint. If titrated, the control should be serially diluted in PBS. When the control has been tested for the endpoint titer by Bion, an endpoint titer is printed on the positive control vial. Due to variations within each laboratory (fluorescent microscope, etc.), each laboratory should establish its own mean titer for each lot of positive control (generally $\pm$ one dilution from stated endpoint).

NEGATIVE CONTROL SERUM

One vial containing 0.5 ml EBV-VCA negative human control serum with protein stabilizer and 0.005% thimerosal. Stable at 2-8°C until labeled expiration date.

The control is intended to be used undiluted as provided. The staining should exhibit less than 1+ fluorescence.

MOUNTING MEDIUM

One dropper vial containing 3.5 ml phosphate buffered glycerol of pH 7.4 ± 0.2 . Stable at 2-8°C until labeled expiration date.

FLUORESCENT ANTIBODY CONJUGATE

Two ready to use dropper vials, each containing 3.5 ml fluorescein isothiocyanate labeled goat antihuman IgG (heavy chain specific) with 0.1% Evans Blue counterstain, protein stabilizer, less than 0.1% sodium azide and 0.001% thimerosal added. Stable at 2-8°C away from direct light until labeled expiration date.

PHOSPHATE BUFFERED SALINE (PBS)

Two one-liter packets of dry PBS. Stable in sealed packet at 25°C, or lower, until labeled expiration date.

BUFFER PREPARATION

Place contents of a one-liter PBS packet in a one-liter volumetric flask, add *distilled water to the one-liter mark, mix and leave several hours or overnight to dissolve. Reconstituted buffer should have a pH of 7.4 ± 0.2 . Adjust with 1N NaOH or 1N HCL if pH value is outside the stated range. Store in a clean screw capped bottle at 25°C or lower. Stable until labeled expiration date provided no gross contamination is seen. Do not use if pH changes, if the solution turns cloudy, or if a precipitate forms.

* Use deionized water with caution, as pH of this type of water may vary causing the pH of PBS to become unstable upon prolonged storage.

WARNINGS AND PRECAUTIONS

1. For *in vitro* diagnostic use. Thus, only staff trained in methods of *in vitro* diagnostics may perform the test.
2. Substrate slides are for single use only and must not be used more than once.
3. Care should be taken when handling substrate slides due to sharp edges.
4. The antigenic substrates have been fixed in acetone and contain no detectable live Epstein-Barr Virus. However, they should be handled and disposed of as any potentially biohazardous laboratory material.
5. Do not remove slides from pouches until ready for testing. Do not use if pouch has been punctured, as indicated by a flat pouch.
6. All reagents should be brought to room temperature (20-25°C) prior to use.
7. Abnormal test results may be seen if the antigen substrate slides are allowed to dry during the staining procedure.
8. Refrigeration (2-8°C) of kit immediately upon arrival will insure stability until labeled expiration date.
9. Reagents should not be used beyond stated expiration date.
10. Substitution of components other than those provided may yield inconsistent results.
11. Do not expose conjugate to strong light during storage or use.
12. Avoid microbial contamination of all reagents involved in the testing procedure or incorrect results may occur.
13. Incubation times or temperatures other than those specified may give erroneous results.
14. Reusable glassware must be washed and thoroughly rinsed free of detergents.
15. Care should be taken to avoid splashing or generation of aerosols.
16. Previously frozen specimens after thawing should be thoroughly mixed prior to testing. It is recommended that sera freeze thawed no more than one time. If repeated testing is required, it is suggested that specimen be aliquoted.
17. Patient samples, as well as all materials coming into contact with them, should be handled at the Biosafety Level 2 as recommended for any potentially infectious human serum or blood specimen in the CDC/NIH manual "Biosafety in Microbiological and Biomedical Laboratories", 1984 Edition. Never pipette by mouth. Avoid contact with skin and mucous membranes.

WARNINGS AND PRECAUTIONS (continued)

16. Sera used to prepare positive and negative controls have been tested by an FDA approved method and found to be negative (were not repeatedly reactive) for the presence of Hepatitis B surface Antigen (HBsAg) and antibodies to Hepatitis C (HepCAb) and HIV 1 & 2. However, because no test method can offer complete assurance of the absence to these or other infectious agents, these reagents should be handled at the Biosafety Level 2 as recommended for any potentially infectious human serum or blood specimen in the CDC/NIH manual "Biosafety in Microbiological and Biomedical Laboratories," 1984 Edition.
17. The preservatives used in conjugates and controls are toxic if ingested. Azides may react with copper or lead plumbing to form explosive metal azides. When disposing, flush drains with water to minimize build-up of azide and metal compounds.

SPECIMEN COLLECTION

Blood should be collected fasting or at least one hour after meals to avoid lipemic serum, as excess lipids may produce a "film" over the substrate. Aseptically collect 5-8 ml of blood by venipuncture. Allow the blood to clot at room temperature (20-25°C) before separating serum to avoid hemolysis which could interfere with test results. Specimens should be stored refrigerated at 2-8°C and tested within one week of collection. Long term storage should be at -20°C in aliquots to avoid repeated freezing and thawing. Do not store in self-defrosting freezer.

Avoid using contaminated sera as they may contain proteolytic enzymes which will digest the substrate. It is unnecessary to

heat inactivate serum specimens prior to testing; however, sera that have been heat inactivated may be used.

When testing paired samples to look for evidence of recent infection, the acute specimen should be obtained as soon as possible after onset of illness and the convalescent specimen obtained 7-14 days later. Acute and convalescent specimens must be tested simultaneously, in the same assay, looking for a significant change in antibody titer between the paired sera. If the first specimen is obtained too late during the course of the infection, a significant rise in the antibody titer may not be detected.

PROCEDURE

MATERIALS PROVIDED

1. EBV-VCA Antigen Substrate Slides
2. Fluorescent Antibody Conjugate
3. Positive Control Serum
4. Negative Control Serum
5. Phosphate Buffered Saline (PBS)
6. Mounting Medium

MATERIALS REQUIRED BUT NOT PROVIDED

1. One liter volumetric flask or one liter graduated cylinder
2. Distilled water - CAP Type one or equivalent
3. One-liter screw capped container
4. Disposable test tubes (12 x 75 mm or comparable) and rack
5. Disposable serological pipettes
6. Calibrated pipettes to deliver 50 µl, 100 µl and 200 µl, with disposable pipette tips
7. Pasteur pipettes and bulbs
8. Moist chambers
9. Plastic squeeze wash bottle
10. Coplin jars or staining dishes with slide racks
11. 24 x 60 mm #1 coverslips
12. Felt tip marking pen
13. Fluorescence microscope equipped with a mercury or tungsten-halogen light source, a 390-490nm excitation filter and 515-520nm barrier filter, and optics to give a total magnification of 200X or 250X. The excitation wavelength of FITC is 490nm and the emission wavelength is 520nm.

TEST PROCEDURE

1. SPECIMEN PREPARATION

Screening:

Prepare a 1:10 dilution of each patient's serum by adding 0.05 ml (50 µl) of patient's serum to 0.45 ml of PBS.

Semi-quantitation:

Serum dilutions are utilized to measure antibody titer. Each laboratory should establish its own titrating protocol. The following fourfold serial titration is suggested:

- a. Prepare a 1:10 dilution of each patient's serum by adding 0.05 ml (50 µl) of patient's serum to 0.45 ml of PBS in tube #1.
- b. Add 0.3 ml PBS to tubes #2, #3, #4 and #5.

- c. Using a 100 µl pipette, transfer 0.1 ml (100 µl) from tube #1 to tube #2. Mix. Using a new tip for each dilution, transfer 0.1 ml (100 µl) from the second tube to the third, from the third tube to the fourth, and from the fourth tube to the fifth, mixing after each transfer. This will give a fourfold titration with the following dilutions:

Tube #1 = 1:10
Tube #2 = 1:40
Tube #3 = 1:160
Tube #4 = 1:640
Tube #5 = 1:2560

TEST PROCEDURE (continued)

2. SLIDE PREPARATION

Remove reagents and as many slides as are required from the refrigerator or freezer and allow to equilibrate to room temperature (20-25°C) for at least five minutes. Remove slides from sealed foil pouches being careful not to touch the antigen surface. Identify each slide using a felt tip marking pen.

3. SPECIMEN APPLICATION

Using separate Pasteur pipettes, apply one drop (20-30 µl) of the Positive Control, one drop (20-30 µl) of the Negative Control and one drop (20-30 µl) of each patient serum dilution to individual wells of the slide. Do not touch the antigen surface with the pipette while dropping. Do not allow drops to mix, as cross contamination of samples between wells could cause erroneous results.

4. INCUBATION 1

Incubate in a moist chamber at room temperature (20-25°C) for 30 minutes.

NOTE: THE ANTIGEN MUST NOT BE ALLOWED TO DRY DURING ANY OF THE FOLLOWING STEPS. Nonspecific binding may occur if the reagent is allowed to dry on the slide.

5. RINSE 1

Remove slides from moist chamber and rinse GENTLY with PBS using a squeeze wash bottle. Do not focus the PBS stream directly onto the wells. To prevent cross contamination tilt slide first toward wells 1-6 and, running a PBS stream along the midline of the slide, allow the PBS to run off the top edge of the slide. Then, tilt the slide toward wells 7-12 and repeat this procedure, allowing the PBS to run off the bottom edge of the slide.

6. WASH 1

Place slides in Coplin jars or staining dishes and wash in two changes of PBS for not less than five minutes or more than ten minutes each, agitating gently at entry and prior to removal.

7. CONJUGATE APPLICATION

Remove slides from the wash one at a time, shake off excess PBS, dry around outside edges if necessary and return each slide to the moist chamber. Apply one drop of conjugate to each well of each slide, making sure each well is completely covered.

8. INCUBATION 2

Incubate in a moist chamber at room temperature (20-25°C) for 30 minutes. Protect slides from excessive light.

9. RINSE 2

Remove slides from moist chamber and rinse GENTLY with PBS using a squeeze wash bottle. As suggested in step 5., do not focus PBS stream directly onto the wells.

10. WASH 2

Place slides in Coplin jars or staining dishes and wash in two changes of PBS for not less than five minutes or more than ten minutes each, agitating gently at entry and prior to removal.

11. COVERSIP

Remove slides one at a time from last PBS wash, shake off excess PBS and immediately add two to four drops of mounting medium across the slide. Tilt slide and rest the edge of the coverslip against the bottom of the slide allowing the mounting medium to form a continuous bead between the coverslip and slide. Gently lower the coverslip from the bottom of the slide to the top, being careful to avoid air bubbles. Drain excess mounting medium by holding the edge of the slide against absorbent paper. Wipe off back of slide.

12. READ

Examine stained slides as soon as possible using a properly equipped fluorescence microscope. It is recommended that slides be examined on the same day they are stained. If any delay is anticipated, store slides in the refrigerator (2-8°C) away from direct light and read the following day. Do not allow mounting medium to dry between slide and coverslip. If drying should occur, add additional mounting medium or recoverslip slide.

FLUORESCENT INTENSITY GRADING

Fluorescent intensity may be semi-quantitated by following the guidelines established by the Centers for Disease Control, Atlanta, Georgia:¹⁴

- 4+ = Maximal fluorescence; brilliant yellow-green.
- 3+ = Less brilliant yellow-green fluorescence.
- 2+ = Definite but dull yellow-green fluorescence.
- 1+ = Very dim subdued fluorescence.

The degree of fluorescent intensity is not clinically relevant and has only limited value as an indicator of titer. Differences in fluorescence microscope optics, filters and light sources may result in differences of 1+ or more fluorescent intensity when observing the same slide using different microscopes.

QUALITY CONTROL

SPECIFICITY CONTROL

Both a positive and negative antibody control must be included with each run. These controls must be examined prior to reading test samples and should demonstrate the following results:

Negative Control

Using the Bion NEGATIVE EBV-VCA CONTROL SERUM as provided with the Bion EBV-G VCA ANTIBODY TEST SYSTEM, the infected cells should exhibit less than 1+ fluorescence and appear reddish-orange due to the counterstain.

Positive Control

Using the Bion EBV-VCA POSITIVE IgG CONTROL SERUM as provided with the Bion EBV-G VCA ANTIBODY TEST SYSTEM, EBV infected cells should exhibit solid fluorescent staining of the entire cell at an intensity of 3+ or greater. Approximately 10% of the cells should exhibit this specific staining pattern with the uninfected cells staining reddish-orange due to the counterstain.

Each control must demonstrate the expected reaction in order to validate the test. If the controls fail to appear as described above, the test results should not be reported and the test should be repeated. If upon repeat testing the controls still fail to show the proper reaction, do not report test results.

SENSITIVITY CONTROL

A titrated control included with each run tests substrate sensitivity, as well as, checks technique, conjugate quality and the microscope optical system. The endpoint titer of each lot of Bion EBV-VCA POSITIVE IgG CONTROL SERUM must be determined. There must not be more than a twofold difference (+/-) in titer from the stated endpoint. Each run should include the endpoint dilution, one fourfold dilution above and one fourfold dilution below the endpoint dilution. The more concentrated dilution should be positive and the less concentrated dilution negative. If the control does not behave as described, the test results are invalid and the tests should be repeated. If the control again fails to show the proper reaction upon repeat testing, do not report the test results.

READING OF TEST RESULTS

NEGATIVE

A serum dilution is considered negative for EBV-VCA IgG antibodies if the cells exhibit less than 1+ fluorescence and appear reddish-orange due to the counterstain, or if the fluorescence observed is not the specific staining pattern of EBV-VCA.

A sample is considered negative for EBV-VCA IgG antibodies if it exhibits less than 1+ fluorescence at a serum dilution of 1:10 and all greater dilutions, or if the fluorescence observed is not the specific staining pattern of EBV-VCA.

- ... Negative samples may exhibit fluorescent staining of the infected cells slightly greater than the Negative Control, but less than 1+.
- ... Nonspecific staining of all cells observed in some sera at low dilutions is most likely due to the presence of autoantibodies against cellular components in either the nucleus or cytoplasm.
- ... Staining of areas other than the viral infected cells should be interpreted as negative and attention should be directed to specific steps in the staining method (e.g., RINSE and WASH steps).

POSITIVE

A serum dilution is considered positive for EBV-VCA IgG antibodies if, at an intensity of 1+ or greater, there is well defined specific fluorescent staining in the EBV-VCA infected cells. The EBV-VCA fluorescent staining pattern consists of solid staining of the entire cell. The number of cells exhibiting a positive staining reaction and the type of fluorescent staining pattern should closely approximate that seen in the Positive Control.

A sample is considered positive for EBV-VCA IgG antibodies if it exhibits the characteristic EBV-VCA staining pattern with a fluorescent intensity of 1+ or greater at a serum dilution of 1:10 or greater.

NOTES: 1. Each field should contain cells that exhibit no apple-green fluorescence. Should most of the cells in the patient test wells fluoresce apple-green in the nucleus and/or cytoplasm, an autoimmune staining reaction due to the presence of autoantibodies should be considered.^{15,16} It is recommended that such samples be diluted beyond the interference for better interpretation. It is possible that autoantibody staining may mask specific staining such that an interpretation cannot be made. Should this occur, test results should be reported as "Unable to interpret due to the presence of interfering antibodies."

2. In a few sera with high concentrations of IgM, a nonspecific fluorescence of all cells may be noticed and mask specific IgG staining.¹⁷ Samples may be diluted beyond this interference for better interpretation.

TITRATION

If a semi-quantitative titration is performed, the result should be reported as the reciprocal of the last dilution in which 1+ apple-green fluorescent intensity of the specific staining pattern is detected. When reading fourfold serial dilutions, endpoints may be extrapolated where necessary.

EXAMPLE OF ENDPOINT EXTRAPOLATION:

1:10 = 4+
1:40 = 3+
1:160 = 2+
1:640 = +/-

The extrapolated endpoint is reported as 320.

TROUBLESHOOTING

Possible solutions to problems that may occur in immunofluorescent assays are discussed in an accompanying brochure entitled "TROUBLESHOOTING IN IMMUNOFLUORESCENCE".

INTERPRETATION OF RESULTS

RESULT	INTERPRETATION OF SINGLE SAMPLE RESULTS
Less than 10	Negative - Indicates no previous infection with EBV and susceptibility to this agent. <u>NOTE:</u> This may represent a primary infection with the humoral immune response not yet developed to detectable levels. If infection with EBV is still suspected, a second specimen should be obtained 7-14 days later, and the paired specimens tested simultaneously, looking for a seroconversion.
10 or Greater	Positive for EBV antibodies. This may represent: 1. a primary infection, reinfection, or reactivation of latent virus; 2. a past experience with EBV; 3. a passively acquired antibody from recent blood transfusions, organ transplantation, or transplacental transfer.

- NOTE:
1. The EBV-VCA IgG titer of a single specimen should not be the only criteria used to aid in the diagnosis of Infectious Mononucleosis. Paired samples (acute and convalescent) must be collected and tested simultaneously in the same assay to look for a seroconversion or a significant rise in titer.
 2. Testing for Early Antigen (Diffuse and Restricted) and IgM specific EBV-VCA antibodies may help confirm a diagnosis of primary infection when only one specimen is available.
 3. Testing for antibodies to EB Nuclear Antigen may help to differentiate a primary infection from either a previous infection or a reactivation of latent virus.

ACUTE RESULT	CONVALESCENT RESULT	INTERPRETATION OF PAIRED SAMPLE RESULTS
Less than 10	Less than 10	Not likely to be an acute EBV infection. <u>NOTE:</u> This may represent a primary infection if time of obtaining the second specimen is too soon after the first. If this condition is suspected, obtain a third specimen 7-14 days after the second specimen and run the three simultaneously, looking for a seroconversion.
Less than 10	10 or Greater	Most likely a primary infection with EBV unless the individual has recently acquired passive antibody.
10 or Greater	10 or Greater but with less than a fourfold difference in titer from the acute specimen	Not significant evidence of current infection. Most likely a previous experience with EBV. This may represent: 1. a primary infection, but the interval between the first and second specimens may not have been long enough for development of a fourfold rise in antibody titer. If this condition is suspected, obtain a third specimen 7-14 days after the second specimen and run the three simultaneously looking for a significant rise in antibody titer; 2. a primary infection, reinfection, or reactivation of latent virus if first specimen is obtained late after onset and antibodies have already reached a plateau; 3. a passively acquired antibody from transplacental transfer, blood transfusion, etc. <u>NOTE:</u> Testing for IgM specific EBV antibodies may help to confirm a diagnosis of active EBV infection when there is less than a fourfold difference in titer between the acute and convalescent specimens.
10 or Greater	10 or Greater with a fourfold or more difference in titer from the acute specimen	Usually indicates an active or recently active EBV infection, be it a primary infection, reinfection or reactivation of latent virus.

INTERPRETATION OF RESULTS (continued)

SEROLOGICAL RESPONSES TO EPSTEIN-BARR VIRUS INFECTION ¹⁰						
PB ^a	VCA-IgM	VCA-IgG	EA/D ^b	EA/R ^c	EBNA ^d	INTERPRETATION
0	0	0	0	0	0	Susceptible
+ or 0 ¹⁸	+	+	+ or 0	+ or 0	0	Acute Primary
0	+ or 0 ^{19,20}	+	+ or 0	+ or 0	+ or 0 ^{10,20}	Recent Primary
0	0	+	0	0	+	Past Infection
0	0	+	+ or 0	+ or 0	+	Past Infection with Possible Reactivation
0	0	++	0	++	+	Burkitt's Lymphoma
0	0	++	++	+ or 0	++	NPC

Key: a = Paul Bunnell Heterophile Antigen
b = Early Antigen, Diffuse

c = Early Antigen Restricted
d = Epstein-Barr Nuclear Antigen

LIMITATIONS OF THE PROCEDURE

- EBV antibody test results should be used in conjunction with information available from clinical evaluation and other diagnostic information.
- A single serological IgG antibody titer to EBV-VCA should not be used as the only criterion for diagnosis of Infectious Mononucleosis. Paired serum samples (acute and convalescent) and testing for IgM specific EBV-VCA antibodies may provide more meaningful data.
- A negative test result does not necessarily rule out current or recent infection. The specimen may have been collected too early in the disease before demonstrable antibody is present.
- Lack of significant rise in titer does not exclude the possibility of recent infection but may indicate an acute phase specimen was obtained too late.
- Positive test results from cord blood or neonates should be interpreted with caution. The presence of IgG antibody in cord blood is usually the result of passive transfer from mother to the fetus. A negative test, however, may be useful in excluding possible infection.²¹
- Test results on specimens from immunosuppressed patients and pregnant women may be difficult to interpret.
- Positive test results may not be valid in persons who have received blood transfusions or various blood products within the past several months.
- Antinuclear antibodies (ANA) present in serum may interfere with the EBV-VCA IFA test. They can be differentiated from EBV staining in that ANAs stain the nuclei in all cells; whereas, EBV-VCA antibodies exhibit staining only in the 10% infected cells.¹⁵
- Cytoplasmic fluorescence in the majority of the cells may be due to the presence of antimitochondrial antibodies (AMA) often seen in primary biliary cirrhosis.¹⁶ They can be differentiated from the specific antigen staining in that AMA will stain the cytoplasm of all cells; whereas, EBV-VCA antibodies exhibit staining only in the 10% infected cells.
- Endpoint reactions may vary between laboratories due to differences in type or condition of fluorescence microscope employed, diluting apparatus, as well as the experience level of personnel performing the assay.
- If both the positive and negative control substrate cells are not visible when viewed using the fluorescence microscope, it may be necessary to replace or realign the light source and check the specific filters.
- Cell culture substrate slides may exhibit nonspecific fluorescence due to contamination of antibodies or PBS rinse-wash solutions with bacteria or fungi. It is very important that personnel reading the staining results have experience in fluorescence microscopy.

SPECIFIC LIMITATIONS OF THE EBV-VCA ASSAY

A false negative reaction may result due to a prozone reaction particularly if screening at only one low dilution such as 1:10.^{22,23}

EXPECTED VALUES

Normally, in classical Infectious Mononucleosis (IM), EBV-VCA antibody titers develop early, reach peak titers within two to four weeks and then decline to lower levels which persist indefinitely.²⁴ A fourfold increase in titer between acute and convalescent sera is diagnostic of an acute or recent IM infection. A single high titer serum above 1:640 is strongly suggestive of recent IM. The titer of EBV-VCA

antibodies may not reflect the severity of clinical symptoms in IM.¹⁰ Approximately 80-90% of the U.S. adult population is positive for antibody to EBV-VCA.²⁵ A subsequent rise in EBV-VCA antibodies, sometimes in excess of 1:2560 may be the result of secondary disease such as Burkitt's Lymphoma or Nasopharyngeal Carcinoma.^{8,20}

SPECIFIC PERFORMANCE CHARACTERISTICS

To investigate the relative specificity and sensitivity of the Bion EBV-G (VCA) ANTIBODY TEST SYSTEM, one hundred specimens were compared qualitatively and sixteen serum specimens were compared semi-quantitatively with another commercially available indirect fluorescent EBV-VCA Antibody Test System.

As summarized in TABLE 1, there was eventually agreement between the two test systems; however, five specimens on initial qualitative screening alone were positive with the BION kit and negative with the OTHER commercial test kit. When quantitated, all five specimens were positive on both systems. Therefore, screening at only one dilution on some systems may result in a false negative result due to prozone reactions.^{22,23} No prozone reaction problems were demonstrated with the Bion EBV-G (VCA) ANTIBODY TEST SYSTEM.²⁶

TABLE 1 - SUMMARY OF RELATIVE COMPARISON TESTING

				Relative Sensitivity	Relative Specificity
OTHER KIT	BION KIT				
		Positive	Negative	100%	86.50%
	Positive	63(5)*	0	63/63	
	Negative	0(5)*	32		32/37

*When titrated, these 5 specimens were also positive with the other kit system.

SPECIFIC PERFORMANCE CHARACTERISTICS (continued)

In addition, sixteen serum specimens were compared semi-quantitatively on both the BION test system and the other commercial test system to investigate the relative sensitivity of the Bion EBV-G (VCA) ANTIBODY TEST SYSTEM. As summarized in TABLE 2, all sixteen specimens ranging in titer from less than 1:10 to 1:2560 agreed with no more than one twofold difference (+/-) in titer with both systems. The tables represent results from two independent readers. If the readers differed, both results are given.²⁶

TABLE 2 - SUMMARY OF RELATIVE SENSITIVITY TESTING

Spec.#	BION	Other	Spec.#	BION	Other
1	<10	<10	9	160	160
2	<10	<10	10	320	160/320
3	<10	<10	11	320	160/640
4	<10	<10	12	640	320/640
5	10	10	13	640	640
6	10	10	14	640/1280	320/640
7	40	20/40	15	640/1280	1280
8	80/160	80	16	2560	1280/2560

Interlot and intralot precision of the Bion EBV-G (VCA) ANTIBODY TEST SYSTEM was evaluated by testing ten serum specimens (3 negative and 7 positive over a range of titers) using four different lot numbers of slides, as well as, testing them three times on slides from the same lot. There was no more than a twofold difference (+/-) in titer between any of the comparison testings, which is within the confidence limits of this methodology. None of the tests vacillated between a positive or a negative result. This data is summarized in TABLE 3 and TABLE 4.²⁶

TABLE 3 - SUMMARY ON INTERLOT PRECISION

Spec.#	Lot 1	Lot 2	Lot 3	Lot 4
1	<10	<10	<10	<10
2	<10	<10	<10	<10
3	<10	<10	<10	<10
4	10	10	20/10	10
5	10	10	10	10
6	80/160	80/160	160	160
7	160	160	160/320	160
8	160/320	320	320	320
9	640	640	640	640
10	640	640	640	640

TABLE 4 - SUMMARY OF INTRALOT PRECISION

Spec.#	1st Test	2nd Test	3rd Test
1	<10	<10	<10
2	<10	<10	<10
3	<10	<10	<10
4	10	10	10
5	10	10	10
6	160	160	160
7	320	320	320
8	640	640	640
9	640/1280	640	640/1280
10	1280	640	640/1280

As Epstein-Barr Virus is a member of the Herpesvirus Group, a study was performed to insure there were no cross-reactions between antibodies to the other members of this group and the Epstein-Barr antigen on the Bion EBV-G (VCA) ANTIBODY TEST SYSTEM substrate slides. Eight serum specimens with IgG antibodies to Cytomegalovirus (CMV), Herpes Simplex Virus Type 1 (HSV 1), Herpes Simplex Virus Type 2 (HSV 2) and/or Varicella Zoster Virus (VZV) were tested using the Bion EBV-G (VCA) ANTIBODY TEST SYSTEM. All eight specimens were negative on the Bion EBV-VCA antigen substrate; thus, there was no cross-reactivity with the EBV substrate and antibodies to other members of this group. Summary of this data is presented in TABLE 5.²⁶

TABLE 5 - SUMMARY OF HERPES GROUP SPECIFICITY STUDY

Spec.#	EBV	CMV	HSV 1	HSV 2	VZV
1	<10	20/40	<10	<10	<10
2	<10	<10	160/320	80/160	20
3	<10	<10	160/320	80	20/40
4	<10	40	<10	<10	<10
5	<10	<10	640	320	20/40
6	<10	<10	320	160	40
7	<10	10	<10	<10	20
8	<10	<10	<10	<10	40/80

BIBLIOGRAPHY

- Epstein, M.A., B.G. Achong, Y.M. Barr, Virus Particles in Cultured Lymphoblasts from Burkitt's Lymphoma, *Lancet*, 1:702-703, 1964.
- Epstein, M.A., Y.M. Barr, B.G. Achong, Studies with Burkitt's Lymphoma, *Wistar Inst. Sympos. Monogr.*, 4:69-82, 1965.
- Henle, G., W. Henle, V. Diehl, Relation of Burkitt's Tumor-Associated Herpes-type Virus to Infectious Mononucleosis, *Proc. Nat. Acad. Sci.*, 59:94-101, 1968.
- Henle, W., G. Henle, Epstein-Barr Virus: The Cause of Infectious Mononucleosis, A Review in: *Oncogenesis and Herpesviruses*, Biggs, I.M., G. de The, L.N. Payne (eds), IARC, Sci. Publ. No.2, Lyon, 269-274, 1972.
- Miller, D., Epstein-Barr Herpes Virus and Infectious Mononucleosis, *Prog. Med. Virol.*, 20:84-112, 1975.
- Henle, W., G. Henle, C.A. Horwitz, Epstein-Barr Virus Specific Diagnostic Tests in Infectious Mononucleosis, *Human Path.*, 5:551-564, 1974.
- Henle, W., G. Henle, Epstein-Barr Virus-Related Serology in Hodgkin's Disease, *Natl. Cancer Inst. Monogr.*, 36:79-84, 1973.
- Evans, A.S., The Spectrum of Infections with EB Virus: A Hypothesis, *J. Infect. Dis.*, 124:330-335, 1971.
- Takada, M., et al., The Establishment of Cultured Cell Lines from the Patients with Iizumi Fever and Infectious Mononucleosis Like Diseases, *Japan J. Exp. Med.*, 43:209-214, 1973.
- Lenette, E.T., Diagnosis of Epstein-Barr Virus Infections, in: *Laboratory Diagnosis of Viral Infections*, Lenette, E.H. (ed), Marcel Dekker, New York, 1985.
- Henle, G., W. Henle, The Diagnosis of Epstein-Barr Virus Infections, in: *The Human Herpesviruses*, Nahamias, A.J., W.R. Dowdle, R.F. Schinazi (eds), Elsevier, New York, 374-378, 1981.
- Weller, T.H., A.H. Coons, Fluorescent Antibody Studies with Agents of Varicella and Herpes Zoster Propagated In Vitro, *Proc. Soc. Exp. Biol. Med.*, 86:789-794, 1954.
- Riggs, J.L., R.J. Siewald, J.H. Burckhalter, C.M. Downs, T.G. Metcalf, Isothiocyanate Compounds as Fluorescent Labeling Agents for Immune Serum, *Am. J. Pathol.*, 34:1081-1097, 1958.
- Lyerla, H.C., F.T. Forrester, The Immunofluorescence (IF) test, in: *Immunofluorescence Methods in Virology*, USDHHS, Georgia, 71-81, 1979.
- Holborow, E.J., D.M. Weir, G.D. Johnson, A Serum Factor in Lupus Erythematosus With Affinity for Tissue Nuclei, *Br. Med. J.*, 11:732-734, 1957.
- Berg, P.A., I. Roitt, D. Doniach, H.M. Cooper, Mitochondrial Antibodies in Primary Biliary Cirrhosis, *Immunol.*, 17:281-293, 1969.
- Schmitz, H. and M. Scherer, IgM Antibodies to Epstein-Barr Virus in Infectious Mononucleosis, *Arch. Gesmt Virusforsch.*, 37:332-339, 1972.
- McCormack, J.G., E. Nowakowski, B.B. Fernandez, The Place of Epstein-Barr Virus Testing in a Clinical Laboratory, *Lab. Med.*, Vol. 8 No. 8:12-15, 1977.
- Smith, T., Director, Dept. of Virology, Section of Clinical Microbiology, Mayo Clinic, Personal Communication.
- Goldman, J.M., M.L. Goodman, D. Miller, Antibody to EB Virus in American Patients with Carcinoma of the Nasopharynx, *J.A.M.A.*, 216:1618-1622, 1971.
- Chernesky, M.A., C.G. Ray, T.F. Smith, Laboratory Diagnosis of Viral Infections, *Cumitech 15*, ASM, Washington, D.C., March 1982.
- Personal communications, Ascher, M., Manager, Immunology, Virology & Microbiology Laboratories, Elmhurst Memorial Hospital, Elmhurst, Illinois.
- Personal Communications, Paxton, H., Director, Immunology, Maryland Medical Laboratories, Baltimore, Maryland.
- Neiderman, J.C., et al., Infectious Mononucleosis: Clinical Manifestations in Relation to EB Virus Antibodies, *J.A.M.A.* 203:205-209, 1968.
- Tischendorf, P., et al., Development and Persistence of Immunity to Epstein-Barr Virus in Man, *J. of Infect. Dis.*, 122:401-409, 1970.
- Data on file, Bion, Des Plaines, Illinois.

AESKU.BION
AN AESKU.GROUP COMPANY



Bion Enterprises, Ltd.
455 State Street, Suite 100
Des Plaines, IL 60016 USA
Phone: +1-847-544-5044
Fax: +1-847-544-5051

EC REP

QbD RepS BV,
Groenenborgerlaan 16,
2610 Wilrijk, Belgium