




AESKU.DIAGNOSTICS
THE DIAGNOSTIC TOOL THAT WORKS



AESKULISA[®]
THE DIAGNOSTIC TOOL THAT WORKS

INSTRUCTION MANUAL

AESKULISA[®] MPO

REF 30-7303US





Product Ref.	30-7303US
Product Desc.	MPO
Manual Rev. No.	002: 2026-01-28

Instruction manual

Table of Contents

1	Intended Use	1
2	Clinical Application and Principle of the Assay	1
3	Kit Contents	2
4	Storage and Shelf Life	2
5	Precautions of Use	3
6	Sample Collection, Handling and Storage	3
7	Assay Procedure	4
8	Semiquantitative and Qualitative Interpretation	5
9	Technical Data	6
10	Performance Data	6
11	Literature	9
12	Regulatory Symbols	10
Annex		11



Aesku.Diagnostics GmbH & Co. KG
Mikroforum Ring 2
55234 Wendelsheim
Germany
Phone: +49 6734 9622-0
Fax: +49 6734 9622-2222
Website: www.aesku.com
Mail: Info@aesku.com

1 Intended Use

AESKULISA® MPO is a solid phase enzyme immunoassay employing purified native myeloperoxidase (MPO) from human peripheral blood polymorphonuclear cells for the semiquantitative and qualitative detection of antibodies against MPO in human serum.

The assay is an aid in the diagnosis of autoimmune systemic vasculitis such as microscopic polyangiitis, and glomerulonephritis and should be used in conjunction with other laboratory and clinical findings.

2 Clinical Application and Principle of the Assay

Antibodies against MPO belong to the group of anti-neutrophil cytoplasmic antibodies (ANCA) which are directed against cytoplasmic components of neutrophilic granulocytes and monocytes¹.

Indirect immunofluorescence tests (IFT) on ethanol-fixed neutrophils have been the established method for the detection of ANCAs. It became apparent that some ANCAs create a cytoplasmic fluorescence pattern (thus called cANCA) while others create a perinuclear pattern (the pANCA).^{2,3}

As both patterns may cover multiple antigens, immunofluorescence is not suitable for a satisfying differential diagnosis of vasculitis; thus, each IFT should be verified with specific ELISA tests.⁴

Whereas proteinase 3 is the main antigen specific for cANCA, the main target antigen for pANCA has been identified as MPO. However, other cellular components (lactoferrin, cathepsin G, elastase etc.) may also cause a perinuclear staining pattern.^{3,5,6}

MPO is an enzyme from the primary granules of neutrophils with a molecular weight of approximately 140 kDa.^{7,8} Its highly positive charge may be relevant for the location at negatively charged structures such as the nuclear membrane thus responsible for the perinuclear staining pattern of anti-MPO antibodies in patients' sera in IFT using ethanol-fixed neutrophils.^{6,9}

ANCAs are important markers for the differential diagnosis of autoimmune vasculitis. Antibodies against MPO are correlated with idiopathic or vasculitis associated necrotizing crescentic glomerulonephritis and are found frequently in about 60% of patients with microscopic polyangiitis, in 10-20% of patients with Wegener's granulomatosis and 30-50% of patients with Churg-Strauss syndrome.^{3,10}

Principle of the test

Serum samples diluted 1:101 are incubated in the microplates coated with the specific antigen. Patient's antibodies, if present in the specimen, bind to the antigen. The unbound fraction is washed off in the following step. Afterwards anti-human immunoglobulins conjugated to horseradish peroxidase (conjugate) are incubated and react with the antigen-antibody complex of the samples in the microplates. Unbound conjugate is washed off in the following step. Addition of TMB-substrate generates an enzymatic colorimetric (blue) reaction, which is stopped by diluted acid (color changes to yellow). The rate of color formation from the chromogen is a function of the amount of conjugate bound to the antigen-antibody complex and this is proportional to the initial concentration of the respective antibodies in the patient sample.



3 Kit Contents

To be reconstituted:

5x Sample Buffer 1 vial, 20 ml - 5x concentrated (capped white: yellow solution)
Containing: Tris, NaCl, BSA, sodium azide <0.1% (preservative)

50x Wash Buffer 1 vial, 20 ml - 50x concentrated (capped white: green solution)
Containing: Tris, NaCl, Tween 20, sodium azide <0.1% (preservative)

Ready to use:

Negative Control 1 vial, 1.5 ml (capped green: colorless solution)
Containing: Human serum (diluted), sodium azide <0.1% (preservative)

Positive Control 1 vial, 1.5 ml (capped red: yellow solution)
Containing: Human serum (diluted), sodium azide <0.1% (preservative)

Cut-off Calibrator 1 vial, 1.5 ml (capped blue: yellow solution)
Containing: Human serum (diluted), sodium azide <0.1% (preservative)

Calibrators 6 vials, 1.5 ml each 0, 3, 10, 30, 100, 300 U/ml
(color increasing with concentration: yellow solutions)
Containing: Human serum (diluted), sodium azide <0.1% (preservative)

Conjugate 1 vial, 15 ml IgG (capped blue: blue solution)
Containing: Goat Anti-human IgG conjugated to horseradish peroxidase, BSA, PBS

TMB Substrate 1 vial, 15 ml (capped black)
Containing: Stabilized TMB/H₂O₂

Stop Solution 1 vial, 15 ml (capped white: colorless solution)
Containing: 1M Hydrochloric Acid

Microtiterplate 12x8 well strips with breakaway microwells
Coating purified native human MPO

Material required but not provided:

Microtiter plate reader 450 nm reading filter and optional 620 nm filter (600-690 nm). Glass ware (cylinder 100-1000 ml), test tubes for dilutions. Vortex mixer, precision pipettes (10, 100, 200, 500, 1000 µl) or adjustable multipipette (100-1000 µl). Microplate washing device (300 µl repeating or multichannel pipette or automated system), adsorbent paper. Our tests are designed to be used with purified water according to the definition of the United States Pharmacopeia and the European Pharmacopeia.

4 Storage and Shelf Life

Store all reagents and the microplate at 2-8°C/35.6-46.4°F, in their original containers. Once prepared, reconstituted solutions are stable for 1 month at 2-8°C/35.6-46.4°F, at least.

Reagents and the microplate shall be used within the expiry date indicated on each component, only. Avoid intense exposure of TMB solution to light. Store microplates in designated foil, including the desiccant, and seal tightly.

5 Precautions of Use

5.1 Health hazard data

THIS PRODUCT IS FOR IN VITRO DIAGNOSTIC USE ONLY. Thus, only staff trained and specially advised in methods of in vitro diagnostics may perform the kit. Although this product is not considered particularly toxic or dangerous in conditions of normal use, refer to the following for maximum safety:

Recommendations and precautions

This kit contains potentially hazardous components. Though the kit reagents are not classified being irritant to eyes and skin we recommend to avoid contact with eyes and skin and wear disposable gloves.

WARNING! Calibrators, Controls and Buffers contain sodium azide (NaN_3) as a preservative. NaN_3 may be toxic if ingested or adsorbed by skin or eyes. NaN_3 may react with lead and copper plumbing to form highly explosive metal azides. On disposal, flush with a large volume of water to prevent azide build-up. Please refer to decontamination procedures as outlined by CDC or other local/national guidelines.

Avoid contact with 1N HCl. It may cause skin irritation and burns. If contact occurs, wash with copious amounts of water and seek medical attention if irritation persists.

Do not smoke, eat or drink when manipulating the kit. Do not pipette by mouth.

All human source material used for some reagents of this kit (controls, standards e.g.) has been tested by FDA approved methods and found negative for HbsAg, Hepatitis C and HIV 1. However, no test can guarantee the absence of viral agents in such material completely. Thus, handle kit controls, standards and patient samples as if capable of transmitting infectious diseases and according to national requirements.

5.2 General directions for use

Do not mix or substitute reagents or microplates from different lot numbers. This may lead to variations in the results.

Allow all components to reach room temperature (20-32°C/68-89.6°F) before use, mix well and follow the recommended incubation scheme for an optimum performance of the test.

Never expose components to higher temperature than 37°C/98.6°F.

Always pipette substrate solution with brand new tips only. Protect this reagent from light. Never pipette conjugate with tips used with other reagents prior.

5.3 Limitations

The test performance has not been established for the pediatric population.

A definite clinical diagnosis should not be based on the results of the performed test only but should be made by the physician after all clinical and laboratory findings have been evaluated. The diagnosis is to be verified using different diagnostic and medicinal methods.

6 Sample Collection, Handling and Storage

Use preferentially freshly collected serum samples. Blood withdrawal must follow national requirements.

Do not use icteric, lipemic, hemolysed or bacterially contaminated samples. Sera with particles should be cleared by low speed centrifugation ($<1000 \times g$). Blood samples should be collected in clean, dry and empty tubes. After separation, the serum samples should be used during the first 8h, respectively stored tightly closed at 2-8°C/35.6-46.4°F up to 48h, or frozen at -20°C/-4°F for longer periods (according CLSI Guideline H18-A3).

7 Assay Procedure

7.1 Preparations prior to pipetting

Dilute concentrated reagents:

Dilute the concentrated sample buffer 1:5 with distilled water (e.g. 20 ml plus 80 ml).

Dilute the concentrated wash buffer 1:50 with distilled water (e.g. 20 ml plus 980 ml).

Samples

Dilute serum samples 1:101 with sample buffer (1x).

e.g. 1000 µl sample buffer (1x) + 10 µl serum. Mix well !

Washing

Prepare 20 ml of diluted wash buffer (1x) per 8 wells or 200 ml for 96 wells

e.g. 4 ml concentrate plus 196 ml distilled water

Automated washing:

Consider excess volumes required for setting up the instrument and dead volume of robot pipette.

Manual washing:

Discard liquid from wells by inverting the plate. Knock the microwell frame with wells downside vigorously on clean adsorbent paper. Pipette 300 µl of diluted wash buffer into each well, wait for 20 seconds. Repeat the whole procedure twice again.

Microplates

Calculate the number of wells required for the test. Remove unused wells from the frame, replace and store in the provided plastic bag, together with desiccant, seal tightly (2-8°C/35.6-46.4°F)

7.2 Work flow

For pipetting scheme see Annex A, for the test procedure see Annex B

We recommend pipetting calibrators in duplicate.

Cut-off calibrator should be used for qualitative testing only.

- Pipette 100 µl of each patient's diluted serum into the designated microwells.
- Pipette 100 µl calibrators OR cut-off control and negative and positive controls into the designated wells.
- Incubate for 30 minutes at room temperature (20-32°C/68-89.6°F).
- Wash 3x with 300 µl washing buffer (diluted 1:50).
- Pipette 100 µl conjugate into each well.
- Incubate for 30 minutes at room temperature (20-32°C/68-89.6°F).
- Wash 3x with 300 µl washing buffer (diluted 1:50).
- Pipette 100 µl TMB substrate into each well.
- Incubate for 30 minutes at room temperature (20-32°C/68-89.6°F).
- Pipette 100 µl stop solution into each well, using the same order as pipetting the substrate.
- Incubate 5 minutes minimum.
- Agitate plate carefully for 5 sec.
- Read absorbance at 450 nm (optionally 450/620 nm) within 30 minutes.

8 Semiquantitative and Qualitative Interpretation

For **semiquantitative interpretation** establish the standard curve by plotting the **optical density (OD) of each calibrator (y-axis)** with respect to the corresponding concentration values in **U/ml (x-axis)**. For best results we recommend log/lin coordinates and 4-Parameter Fit. From the OD of each sample, read the corresponding antibody concentrations expressed in **U/ml**.

Normal Range	Positive Results
≤ 15 U/ml	> 15 U/ml

Example of a standard curve

We recommend pipetting calibrators in parallel for each run.

Calibrators IgG	OD 450/620 nm	CV % (Variation)
0 U/ml	0.055	0.1
3 U/ml	0.195	0.7
10 U/ml	0.400	2.4
30 U/ml	0.785	0.5
100 U/ml	1.440	1.7
300 U/ml	2.300	0.9

Example of calculation

Patient	Replicate (OD)	Mean (OD)	Result (U/ml)
P 01	0.794/0.792	0.793	32.1
P 02	1.345/1.321	1.333	84.5

For lot specific data, see enclosed quality control leaflet. Medical laboratories might perform an inhouse quality control by using own controls and/or internal pooled sera, as foreseen by national regulations.

Do not use this example for interpreting patients results!

Each laboratory should establish its own normal range based upon its own techniques, controls, equipment and patient population according to their own established procedures.

For qualitative interpretation read the optical density of the cut-off calibrator and the patient samples. Compare patient's OD with the OD of the cut-off calibrator. All samples with higher ODs are considered positive, samples with lower ODs are considered negative.

Samples above the calibrator range should be reported as > 300 U/ml. They should be diluted as appropriate and reassayed. Samples below calibrator range should be reported as $< \text{Min}$.

Negative: $\text{OD}_{\text{patient}} \leq \text{OD}_{\text{cut-off}}$

Positive: $\text{OD}_{\text{patient}} > \text{OD}_{\text{cut-off}}$

9 Technical Data

Sample material:	serum
Sample volume:	10 µl of sample diluted 1:101 with 1x sample buffer
Total incubation time:	90 minutes at 20-32°C/68-89.6°F
Calibration range:	0-300U/ml
Analytical sensitivity:	1.47 U/ml
Storage:	at 2-8°C/35.6-46.4°F use original vials, only
Number of determinations:	96 tests

10 Performance Data

10.1 Limit of Detection

Testing sample buffer 60 times on **AESKULISA® MPO** and 8 low negative samples for 8 times gave a limit of detection of 1.47 U/ml.

10.2 Positive and Negative agreement with Predicate device

The microplates are coated with native **human Myeloperoxidase**. Antibodies against MPO are correlated with idiopathic or vasculitis associated necrotizing crescentic glomerulonephritis and are found frequently in about 60% of patients with microscopic polyangiitis, in 10-20% of patients with Wegener's granulomatosis and 30-50% of patients with Churg-Strauss syndrome.^{3,10}

151 sera of patients suffering from Wegener's granulomatosis, microscopic polyangiitis and other autoimmune diseases have been tested on the **AESKULISA® MPO** and a predicate device, of these 79 sera lay in range of the assay and were used for a comparison study versus a predicate device.

		predicate device			
		POS	EQUIV	NEG	TOTAL
AESKULISA® PR3	POS	39	4	0	43
	NEG	0	7	29	36
	TOTAL	39	11	29	79

* An equivocal results of the predicate device has been considered as a negative for this calculation.

		95% C.I.
Overall percent agreement*	94.9 %	87.7% to 98.0%
Positive percent agreement*	100%	91.0% to 100%
Negative percent agreement*	90.0%	77.0% to 96.0%

For a clinical performance study only the samples which should clearly contain MPO antibodies (Glomerulo nephritis, microscopic polyangiitis) were considered as positive for the diagnostic sensitivity/specificity calculation, all other diagnosis, though there may be MPO antibodies present were considered as "to be negative" (complete data upon request).

		95% C.I.
Overall percent agreement*	94.2 %	89.1% to 97.1%
diagnostic sensitivity*	94.1%	80.9% to 98.4%
diagnostic specificity*	94.3%	88.1% to 97.4%

* Only samples with diseases with clear MPO antibody presence have been considered.



		diagnosis		
		POS	NEG	TOTAL
AESKULISA® MPO	POS	32	6	38
	NEG	2	99	101
	Total	34	105	139

Number of Samples with		AESKULISA® MPO		
Diagnosis		POS	NEG	Total
Acute hearing loss		1	0	1
Chronic renal disease		0	1	1
Churg-Strauss		0	2	2
COPD		3	0	3
Crohn's disease		0	6	6
Endocarditis		0	1	1
Goodpasture-Syndrome		0	1	1
healthy		1	0	1
HIV		0	1	1
palsy		0	1	1
Polymyalgia rheumatica (vasculitis)		0	1	1
Reactive Arthritis		0	21	21
Rheumatoid arthritis		0	1	1
SLE		1	0	1
Ulcerative colitis		0	6	6
Ulcerative Colitis (septic fungal infection)		0	1	1
Wegeners Granulomatosis		0	56	56
Glomerulo nephritis (c-ANCA positive) *		0	1	1
Glomerulo nephritis (GN)*		2	0	2
mPAN*		30	0	30
Wegener's Granulomatosis / GN*		0	1	1
Total		38	101	139

10.3 Linearity

Chosen sera have been tested with this kit and found to dilute linearly. However, due to the heterogeneous nature of human autoantibodies there might be samples that do not follow this rule.

Sample No.	Dilution Factor	measured (U/ml)	expected (U/ml)	Recovery (%) 90-110%
1	1 / 100	76.5	78.0	9.1
	1 / 200	37.3	39.0	95.6
	1 / 400	19.2	19.5	98.5
	1 / 800	9.4	9.8	95.9
2	1 / 100	32.8	33.0	99.4
	1 / 200	17.4	16.5	105.5
	1 / 400	9.0	8.3	108.4
	1 / 800	4.2	4.1	102.4
3	1 / 100	342.15	325	105.3
	1 / 200	177.5	162.5	109.2
	1 / 400	85.8	81.25	105.6
	1 / 800	42	40.625	103.4
4	1 / 100	235.5	252	93.5
	1 / 200	121.15	126	96.2
	1 / 400	60.3	63	95.7
	1 / 800	33.65	31.5	106.8



10.4 Precision

To determine the precision of the assay, the variability (intra-, inter-assay and lot-to-lot variability) was assessed by examining its reproducibility on eight serum samples selected to represent a range over the standard curve.

The accepted range for CV is 10% for positive samples, 15% around cut-off and 20% for negative samples (N=40 / 40 / 24).

Intra-Assay			Inter-Assay			Lot-to-Lot variability		
Sample No.	Mean (U/ml)	CV (%)	Sample No.	Mean (U/ml)	CV (%)	Sample No.	Mean (U/ml)	CV (%)
1	6.2	14.3	1	6.2	14.4	1	6.2	12.7
2	7.1	10.6	2	7.1	10.8	2	7.0	10.8
3	10.1	9.0	3	10.1	8.8	3	10.1	8.8
4	14.6	9.4	4	14.6	9.3	4	14.3	9.2
5	25.9	8.0	5	28.9	7.7	5	25.6	6.5
6	38.6	1.6	6	38.6	1.7	6	32.7	3.7
7	78.5	2.5	7	78.5	3.0	7	162.3	7.4
8	173.9	5.7	8	173.9	5.8	8	53.5	8.2

10.5 Calibration

Due to the lack of international reference calibration this assay is calibrated in arbitrary units (U/ml).



11 Literature

- 1. Falk, RJ Jenette JC (1988).** *Antineutrophil cytoplasmic autoantibodies with specificity myeloperoxidase in patients with systemic vasculitis and idiopathic necrotizing and crescentic glomerulonephritis.* N Engl J Med 318:1651-1657.
- 2. Baslund B and Wiik A (1994)** *Anti-neutrophil cytoplasmic autoantibodies (ANCA) and vasculitis.* Clin Rev Allergy; 12: 297-304.
- 3. Bosch X, Guilabert A and Font J (2006).** *Antineutrophil cytoplasmic antibodies.* Lancet; 368: 404-418.
- 4. Savige J, Gillis D, benson et al. (1999).** *International consensus statement on testing and reporting of antineutrophil cytoplasmic antibodies (ANCA).* Am J Clin Pathol; 111: 507-13.
- 5. Talor MV, Stone JH, Stebbing J, Barin J, Rose NR and Burek CL (2007).** *Antibodies to selected minor target antigens in patients with anti-neutrophil cytoplasmic antibodies (ANCA).* Clin Exp Immunol; 150: 42-48
- 6. Wilk A, Stumann L, Kjeldsen L et al. (1995).** *The diversity of perinuclear antineutrophil cytoplasmic antibodies (pANCA) antigens.* Clin Exp Immunol; 101: Suppl 1, 15-17.
- 7. Sutton BJ, Little C, Olsen RL and Willassen NP (1988).** *Preliminary crystallographic analysis of human myeloperoxidase.* J Mol Biol; 199: 395-396.
- 8. Zeng J and Fenna RE (2007).** *X-Ray crystal structure of canine myeloperoxidase at 3 Å resolution.* J Mol Biol; 226: 185-207.
- 9. Radice A and Sinico RA (2005).** *Antineutrophil cytoplasmic antibodies (ANCA).* Autoimmunity; 38: 93-103.
- 10. Kallenberg CGM (2007).** *Antineutrophil cytoplasmic autoantibodies with specificity for myeloperoxidase. In Autoantibodies (ed. Shoenfeld Y, Gershwin ME, Meroni PL) (2007) 2nd Edition 95-103*

12 Regulatory Symbols

IVD	For in vitro diagnostic use
REF	Catalog number
LOT	Lot
UDI	Unique Device Identifier
	96 tests
	See instructions for use
	Use by
	Store at 2-8°C (35.6-46.4°F)
	Manufactured by
CON +	Positive Control
CON -	Negative Control
CAL	Calibrator
CO-CAL	Cut off Calibrator
CONJ	Conjugate
MP	Coated microtiter plate
WASHB 50x	Wash buffer
SUB	Substrate buffer
STOP	Stop solution
SB 5x	Sample buffer
Rx only	For Prescription Use only



Annex

A: Pipetting scheme

We suggest pipetting calibrators, controls and samples as follows:

For **semiquantitative interpretation** use calibrators to establish a standard curve.

For **qualitative interpretation** use cut-off calibrator.

	for semiquantitative interpretation use calibrators to establish a standard curve						for qualitative interpretation use cut-off calibrator					
	1	2	3	4	5	6	7	8	9	10	11	12
A	CalA	CalE	P1				NC	P3				
B	CalA	CalE	P2				NC	P4				
C	CalB	CalF	P3				CC	P5				
D	CalB	CalF	P4				CC	P6				
E	CalC	PC	P5				PC	...				
F	CalC	PC	P6				PC	...				
G	CalD	NC	...				P1	...				
H	CalD	NC	...				P2	...				

CalA: calibrator A, CalB: calibrator B, CalC: calibrator C, CalD: calibrator D, CalE: calibrator E, CalF: calibrator F

PC: positive control

NC: negative control

CC: Cut-off calibrator




P1: patient 1

P2: patient 2

P3: patient 3



B: Test Procedure



C: Testing Protocol

Assay/Test: _____ Incubation/ Inkub. : 1. _____ min Date/Datum: _____
 Temperature/Temperatur: _____ °F _____ °C Signature/Unterschrift: _____
 Name: _____ 3. _____ min

	1	2	3	4	5	6	7	8	9	10	11	12
A												
B												
C												
D												
E												
F												
G												
H												