




AESKU.DIAGNOSTICS
THE DIAGNOSTIC TOOL THAT WORKS



AESKULISA®

THE DIAGNOSTIC TOOL THAT WORKS

INSTRUCTION MANUAL

AESKULISA® Cenp-B

REF 7112US





Product Ref.	7112US
Product Desc.	Cenp-B
Manual Rev. No.	007: 2026-01-09

Instruction manual

Table of Contents

1	Intended Use	1
2	Clinical Application and Principle of the Assay	1
3	Kit Contents	2
4	Storage and Shelf Life.....	2
5	Precautions of Use.....	3
6	Sample Collection, Handling and Storage.....	3
7	Assay Procedure.....	4
8	Quantitative and Qualitative Interpretation	5
9	Technical Data	6
10	Performance Data.....	6
11	Literature.....	7
12	Regulatory Symbols	8
	Annex	9



Aesku.Diagnostics GmbH & Co. KG
Mikroforum Ring 2
55234 Wendelsheim
Germany
Phone: +49 6734 9622-0
Fax: +49 6734 9622-2222
Website: www.aesku.com
Mail: Info@aesku.com

1 Intended Use

AESKULISA® Cenp-B is a solid phase enzyme immunoassay employing purified recombinant human 80 kDa centromere protein B (Cenp-B) for the qualitative and semi-quantitative detection of IgG antibodies against Cenp-B in human serum. The assay serves as an aid in the diagnosis of systemic sclerosis and CREST syndrome and should be used in conjunction with other serological tests and clinical findings.

2 Clinical Application and Principle of the Assay

Antibodies against Scl-70 are directed against DNA-topoisomerase I which is an enzyme located in the nucleoplasm, nucleolus and nucleolar organizing region. It catalyzes the breaking and rejoining of single stranded DNA thus interconverting different topological forms of DNA, most likely during transcription. The whole antigen has a molecular weight of 110 kDa but is easily degraded by proteases to 100 kDa, 87 kDa and 70 kDa (hence Scl-70).

Anti-Scl-70 belong to the heterogenous group of anti-nuclear antibodies (ANA) which are directed against various proteins of the nucleus. ANAs are associated with various autoimmune diseases. Antibodies against Scl-70 as well as antibodies against the chromosomal centromere protein B (Cenp-B) are highly specific for systemic sclerosis, a multisystemic autoimmune disease with systemic fibrosis of the connective tissue. Antibodies against Scl-70 can be found in 70% of patients with diffuse cutaneous systemic sclerosis and give a hint for a severe course. Antibodies against Cenp-B are characteristic for a slower progressing variant of systemic sclerosis, the CREST syndrome (occurrence in 70-80% of CREST patients). The respective antibody has a prognostic value for the disease's progress. It was shown that the presence of both autoantibodies in one patient is very rare.

Principle of the test

Serum samples diluted 1:101 are incubated in the microplates coated with the specific antigen. Patient's antibodies, if present in the specimen, bind to the antigen. The unbound fraction is washed off in the following step. Afterwards anti-human immunoglobulins conjugated to horseradish peroxidase (conjugate) are incubated and react with the antigen-antibody complex of the samples in the microplates. Unbound conjugate is washed off in the following step. Addition of TMB-substrate generates an enzymatic colorimetric (blue) reaction, which is stopped by diluted acid (color changes to yellow). The rate of color formation from the chromogen is a function of the amount of conjugate bound to the antigen-antibody complex and this is proportional to the initial concentration of the respective antibodies in the patient sample.



3 Kit Contents

To be reconstituted:

5x Sample Buffer 1 vial, 20 ml - 5x concentrated (capped white: yellow solution)
Containing: Tris, NaCl, BSA, sodium azide (preservative)

50x Wash Buffer 1 vial, 20 ml - 50x concentrated (capped white: green solution)
Containing: Tris, NaCl, Tween, sodium azide (preservative)

Ready to use:

Negative Control 1 vial, 1.5 ml (capped green: yellow solution)
Containing: Human serum (diluted), sodium azide (preservative)

Positive Control 1 vial, 1.5 ml (capped red: yellow solution)
Containing: Human serum (diluted), sodium azide (preservative)

Cut-off Control 1 vial, 1.5 ml (capped blue: yellow solution)
Containing: Human serum (diluted), sodium Azide (preservative)

Calibrators 6 vials, 1.5 ml each 0, 3, 10, 30, 100, 300 U/ml
(color increasing with concentration: yellow solutions)
Containing: Human serum (diluted), sodium azide (preservative)

Conjugate 1 vial, 15 ml IgG (capped blue: blue solution)
Containing: Anti-human immunoglobulins conjugated to horseradish peroxidase

TMB Substrate 1 vial, 15 ml (capped black)
Containing: Stabilized TMB/H₂O₂

Stop Solution 1 vial, 15 ml (capped white: colorless solution)
Containing: 1M Hydrochloric Acid

Microtiterplate 12x8 well strips with breakaway microwells
Coating see paragraph 1

Material required but not provided:

Microtiter plate reader 450 nm reading filter and optional 620 nm reference filter (600-690 nm). Glass ware (cylinder 100-1000 ml), test tubes for dilutions. Vortex mixer, precision pipettes (10, 100, 200, 500, 1000 µl) or multipipette. Microplate washing device (multichannel pipette or automated system), adsorbent paper. Our tests are designed to be used with purified water according to the definition of the United States Pharmacopeia and the European Pharmacopeia.

4 Storage and Shelf Life

Store all reagents and the microplate at 2-8°C/35.6-46.4°F, in their original containers. Once prepared, reconstituted solutions are stable for 1 month at 2-8°C/35.6-46.4°F, at least.

Reagents and the microplate shall be used within the expiry date indicated on each component, only. Avoid intense exposure of TMB solution to light. Store microplates in designated foil, including the desiccant, and seal tightly.



5 Precautions of Use

5.1 Health hazard data

THIS PRODUCT IS FOR IN VITRO DIAGNOSTIC USE ONLY. Thus, only staff trained and specially advised in methods of in vitro diagnostics may perform the kit. Although this product is not considered particularly toxic or dangerous in conditions of normal use, refer to the following for maximum safety:

Recommendations and precautions

This kit contains potentially hazardous components. Though the kit reagents are not classified being irritant to eyes and skin we recommend to avoid contact with eyes and skin and wear disposable gloves.

Do not smoke, eat or drink when manipulating the kit.

Do not pipette by mouth.

All human source material used for some reagents of this kit (controls, standards e.g.) has been tested by approved methods and found negative for HbsAg, Hepatitis C and HIV 1. However, no test can guarantee the absence of viral agents in such material completely. Thus, handle kit controls, standards and patient samples as if capable of transmitting infectious diseases and according to national requirements.

5.2 General directions for use

Do not mix or substitute reagents or microplates from different lot numbers. This may lead to variations in the results.

Allow all components to reach room temperature (20-26°C/68-78.8°F) before use, mix well and follow the recommended incubation scheme for an optimum performance of the test.

Never expose components to higher temperature than 37°C/ 98.6 °F.

Always pipette substrate solution with brand new tips only. Protect this reagent from light. Never pipette conjugate with tips used with other reagents prior.

A definite clinical diagnosis should not be based on the results of the performed test only, but should be made by the physician after all clinical and laboratory findings have been evaluated. The diagnosis is to be verified using different diagnostic and medicinal methods if the patient has got infectious diseases accompanied by medication.

6 Sample Collection, Handling and Storage

Use preferentially freshly collected serum samples. Blood withdrawal must follow national requirements.

Do not use icteric, lipemic, hemolysed or bacterially contaminated samples. Sera with particles should be cleared by low speed centrifugation (<1000 x g). Blood samples should be collected in clean, dry and empty tubes. After separation, the serum samples should be used immediately, respectively stored tightly closed at 2-8°C/35.6-46.4°F up to three days, or frozen at -20°C/-4°F for longer periods.

7 Assay Procedure

7.1 Preparations prior to pipetting

Dilute concentrated reagents:

Dilute the concentrated sample buffer 1:5 with distilled water (e.g. 20 ml plus 80 ml).

Dilute the concentrated wash buffer 1:50 with distilled water (e.g. 20 ml plus 980 ml).

Samples

Dilute serum samples 1:101 with sample buffer (1x).

e.g. 1000 µl sample buffer (1x) + 10 µl serum. Mix well !

Washing

Prepare 20 ml of diluted wash buffer (1x) per 8 wells or 200 ml for 96 wells

e.g. 4 ml concentrate plus 196 ml distilled water.

Automated washing:

Consider excess volumes required for setting up the instrument and dead volume of robot pipette.

Manual washing:

Discard liquid from wells by inverting the plate. Knock the microwell frame with wells downside vigorously on clean adsorbent paper. Pipette 300 µl of diluted wash buffer into each well, wait for 20 seconds. Repeat the whole procedure twice again.

Microplates

Calculate the number of wells required for the test. Remove unused wells from the frame, replace and store in the provided plastic bag, together with desiccant, seal tightly (2-8°C/35.6-46.4°F).

7.2 Work flow

- Pipette 100 µl of each patient's diluted serum into the designated microwells.
- Pipette 100 µl calibrators OR cut-off control and negative and positive controls into the designated wells.
- Incubate for 30 minutes at room temperature (20-26°C/68-78.8°F).
- Wash 3x with 300 µl washing buffer (diluted 1:50).
- Pipette 100 µl conjugate into each well.
- Incubate for 15 minutes at room temperature (20-26°C/68-78.8°F).
- Wash 3x with 300 µl washing buffer (diluted 1:50).
- Pipette 100 µl TMB substrate into each well.
- Incubate for 15 minutes at room temperature (20-26°C/68-78.8°F), in the dark.
- Pipette 100 µl stop solution into each well, using the same order as pipetting the substrate.
- Incubate 5 minutes minimum.
- Agitate plate carefully for 5 sec.
- Read absorbance at 450 nm (optionally 450/620 nm) within 30 minutes.

8 Quantitative and Qualitative Interpretation

For **quantitative interpretation** establish the standard curve by plotting the **optical density (OD) of each calibrator (y-axis)** with respect to the corresponding concentration values in **U/ml (x-axis)**. For best results we recommend log/lin coordinates and 4-Parameter Fit. From the OD of each sample, read the corresponding antibody concentrations expressed in **U/ml**.

Normal Range	Positive Results
≤ 15 U/ml	> 15 U/ml

Example of a standard curve

We recommend pipetting calibrators in parallel for each run.

Calibrators IgG	OD 450/620 nm	CV %
0 U/ml	0.049	1.9
3 U/ml	0.146	2.1
10 U/ml	0.337	2.7
30 U/ml	0.667	1.3
100 U/ml	1.246	1.6
300 U/ml	2.213	2.3

Example of a calculation

Patient	Replicate (OD)	Mean (OD)	Result (U/ml)
P 01	0.814/0.795	0.805	44.3
P 02	1.568/1.553	1.561	141.7

For lot specific data, see enclosed quality control leaflet. Medical laboratories might perform an in-house Quality Control by using own controls and/or internal pooled sera, as foreseen by EU regulations. **Do not use this example for interpreting patients results!**

Each laboratory should establish its own normal range based upon its own techniques, controls, equipment and patient population according to their own established procedures.

For **qualitative interpretation** read the optical density of the cut-off control and the patient samples. Compare patient's OD with the OD of the cut-off control. All samples which are higher than cut-off are considered positive.

Negative:	OD patient < OD_{cut-off}
Positive:	OD patient > OD_{cut-off}



9 Technical Data

Sample material:	serum
Sample volume:	10 µl of sample diluted 1:101 with 1x sample buffer
Total incubation time:	60 minutes at room temperature (20-26°C/68-78.8°F)
Calibration range:	0-300 U/ml
Analytical sensitivity:	1.0 U/ml
Storage:	at 2-8°C/35.6-46.4°F use original vials, only
Number of determinations:	96 tests

10 Performance Data

10.1 Analytical Sensitivity

The analytical sensitivity of this kit has been found at 1.0 U/ml.

10.2 Specificity and sensitivity

The microplate is coated with **human recombinant 80 kDa centromere protein B**. No cross reactivities to other autoantigens have been found. Up to 80% of CREST syndrome patients and 25% of patients with Raynaud syndrome exhibit centromere antibodies. A study with 50 Anti-CenpB positive and 53 negative sera (from patients with various rheumatic disorders) and the **AESKULISA®** CenpB is shown in the table below.

clinical data for CenpB	results from the Aeskulisa CenpB		
		positive	negative
	positive	50	0
	negative	0	53

100% agreement

10.3 Linearity

Chosen sera have been tested with this kit and found to dilute linearly. However, due to the heterogeneous nature of human autoantibodies there might be samples that do not follow this rule.

Sample No.	Dilution measured Factor	expected concentration (U/ml)	Recovery concentration (U/ml)	(%)
1	1 / 100	271.0	275.0	98.5
	1 / 200	135.0	137.8	98.0
	1 / 400	69.0	68.8	100.3
	1 / 800	34.1	34.4	99.1
2	1 / 100	164.0	166.0	98.8
	1 / 200	84.0	83.0	101.2
	1 / 400	42.1	41.5	101.5
	1 / 800	21.1	20.8	101.4



10.4 Precision

To determine the precision of the assay, the variability (intra and inter-assay) was assessed by examining its reproducibility on three serum samples selected to represent a range over the standard curve.

Intra-Assay			Inter-Assay		
Sample No.	Mean (U/ml)	CV (%)	Sample No.	Mean (U/ml)	CV (%)
1	265.0	1.5	1	261.0	1.4
2	122.0	2.3	2	119.0	2.5
3	44.0	4.1	3	46.0	3.8

10.5 Calibration


The **AESKULISA**® Cenp-B is calibrated against reference sera from the CDC Atlanta (Centers for Disease Control and Prevention). The results are expressed in U/ml.

11 Literature

- Earnshaw, WC, Bordwell, B, Marino, C, Rothfield, N. (1986):**
The three human chromosomal autoantigens are recognized by sera from patients with anticentromere antibodies.
J Clin Invest 77: 426-430
- von Mühlen, CA, Tan, EM (1995):**
Autoantibodies in the diagnosis of systemic rheumatic diseases.
Semin Arthritis Rheum 24: 323-358.
- Tan, EM (1989):**
Antinuclear antibodies: diagnostic markers for autoimmune diseases and probes for cell biology.
Adv Immunol 44: 93-151.
- Fritzler, MJ (1993):**
Autoantibodies in scleroderma.
J Dermatol 24: 323-58.



12 Regulatory Symbols

IVD	For in vitro diagnostic use
REF	Catalog number
LOT	Lot
UDI	Unique Device Identifier
	96 tests
	See instructions for use
	Use by
	Store at 2-8°C (35.6-46.4°F)
	Manufactured by
CON +	Positive Control
CON -	Negative Control
CAL	Calibrator
CO-CAL	Cut off Calibrator
CONJ	Conjugate
MP	Coated microtiter plate
WASHB 50x	Wash buffer
SUB	Substrate buffer
STOP	Stop solution
SB 5x	Sample buffer
Rx only	For Prescription Use only






Annex

A: Pipetting Scheme

		Calibrators A-F	Controls	Samples
Pipette	Calibrators A-F	100 µl each	-	-
Pipette	Controls	-	100 µl each	-
Pipette	Prediluted samples (1:101)	-	-	100 µl each
Incubate	30 min at room temperature (20-26°C/68-78.8°F)			
Decant	Wash 3x with 300 µl of wash buffer (1x)			
Pipette	Conjugate	100 µl	100 µl	100 µl
Incubate	15 min at room temperature (20-26°C/68-78.8°F)			
Decant	Wash 3x with 300 µl of wash buffer (1x)			
Pipette	Substrate	100 µl	100 µl	100 µl
Incubate	15 min at room temperature (20-26°C/68-78.8°F), in the dark.			
Pipette	Stop Solution	100 µl	100 µl	100 µl
Incubate	5 min at room temperature (20-26°C/68-78.8°F)			
Agitate plate for 5 seconds and read OD at λ450nm (optionally λ 450/620 nm) within 30 minutes. Resulting color is stable for 30 minutes, at least.				

B: Test Procedure



C: Test Protocol

Assay/Test: _____		Incubation/ Inkub. : _____		Date/Datum: _____	
Temperature/Temperatur: _____ °F _____ °C		1. _____ min		Signature/Unterschrift: _____	
Name: _____		2. _____ min			
		3. _____ min			

	1	2	3	4	5	6	7	8	9	10	11	12
A												
B												
C												
D												
E												
F												
G												
H												